

MEASUREMENT CHECK LIST

Author: Tuomas Tolvanen

Date: 01-10-2020

The measurement check list is intended to help a new user to start the measurements at AMI Centre.						
☐ Safety	course					
scanner, ar	A new user has to pass AMI Centre's safety course, which includes a video lecture, hands-on demonstration at the scanner, and examination. The safety course is organized several times per year. The safety course and rigorous MR safety procedures has kept us free from any MRI-related medical emergencies since the inauguration of AMI Centre (2002).					
☐ Read A	.MI Centre's Safe	ty and Operations instruct	ions			
A new user	has an obligation	to become acquainted with	AMI Centre's Safety and Opera	ations instructions.		
☐ Resear	ch permission (<u></u>	https://www.aalto.fi/services/a	advanced-magnetic-imaging-ce	entre)		
Obtain a su	pporting stateme	nt from appropriate ethics co	mmittee (e.g. Aalto, HUS, UH,	etc.)		
 Aalto, https://inside.aalto.fi/display/AboutAalto/Research+Ethics+Committee (requires Aalto account) HUS, https://www.hus.fi/tutkijalle/eettiset-toimikunnat/Sivut/default.aspx UH, https://www.helsinki.fi/en/research/research-environment/research-ethics Secure funding List the persons who can reserve scanner They must have passed safety course They must be listed in the research permission as well 						
Name		Email	Phone	Affiliation		
 ■ MRI protocol Select MRI coil (https://www.aalto.fi/en/aalto-neuroimaging-infrastructure-ani/ami-centre-facilities-mri-system) □ 32-channel, best general-purpose research coil □ 30-channel, recommended for eye tracking, very wide binocular field of view □ 20-channel, for simultaneous EEG measurements, slightly larger inner dimensions (note head size) 						
Sa	Select fMRI sequence					
36	•					
•						

https://ani.aalto.fi



MEASUREMENT CHECK LIST

Author: Tuomas Tolvanen

Date: 01-10-2020

fMRI / DTI / MPRAGE etc. parameters

Coil		
Sequence		
Orientation		
TR		
TE		
Slices		
Slice thickness		
Dist. Factor (Gap)		
Phase enc. dir.		
FOV read		
FOV phase		
Averages		
Flip angle		
Fat. Suppr.		
Base resolution		
Phase resolution		
Time points		
PAT (Grappa)		
SMS acc.		
Other parameters		

	O4: 1		
	i Stimiiliie	equipment	r
_	. Othinuus	cquipiliciti	٠

Presentation E-Prime Matlab PsychoPy Psychtoolbox

Software	Version

Stimuli

Are the default settings fine for your experiment? If you have particular requirements for the stimuli, please contact Tuomas Tolvanen (tuomas.j.tolvanen@aalto.fi).

☐ Visual stimuli

Default and recommended settings for visual stimuli

Resolution 1920x1080Frequency 120 Hz

Screen dimensions 33.5 cm x 21.0 cm

Viewing distance 35–39 cm

Resolution	Frequency	Screen dimensions	Viewing distance

NOTE! Large stimulus size -> head movement possible

☐ Auditory stimuli

- Measure dB levels before measurements
- Check default sound volume for your stimuli
 - ☐ Sensimetrics (headphones are not applicable with all sequences, e.g. DTI)

■ ADU headphones

https://ani.aalto.fi 2/4



MEASUREMENT CHECK LIST

Author: Tuomas Tolvanen

Date: 01-10-2020

	(<u>https://www.aa</u>	<u>ılto.fi/en/aalto-r</u>	<u>neuroimaging</u>	<u>-infrastructure</u>	<u>-ani/ami-centre</u>	<u>-facilities-</u>
esponse-devices						

Response devices

Lumitouch – left and/or right hand buttons fORP – left and/or right hand buttons, joystick, trackball, grip force Optical microphone

Device	Handheld

	Mon	itoring	systems
--	-----	---------	---------

□ EEG	☐ Pulse	☐ Face camera (movement/gesture)
□ ECG	☐ Respiratory bellows	☐ Accelerometers
□ GSR	☐ Eve tracking	

□ Piloting

- With a phantom or a human subject
- Free with phantom, contact Toni Auranen (toni.auranen@aalto.fi)
- Human piloting is subject to a charge
- · Check timing with stimuli etc.
- Prepare your MRI imaging protocol
- Develop your own measurement procedure!

☐ Other

- Get Aalto key card from NBE HR Coordinator
- External users please ask instructions from Rami Kunnas (rami.kunnas@aalto.fi) at AMI Centre)
- Get access to MRI from Toni Auranen (toni.auranen@aalto.fi) at AMI Centre

□ Data transfer

- Ask for a AMI user id from Petteri Räisänen (petteri.raisanen@aalto.fi) at the Department of Neuroscience and Biomedical Engineering, Aalto University School of Science
- Ask detailed instructions from Tuomas Tolvanen (tuomas.i.tolvanen@aalto.fi) at AMI Centre.

NOTE! Please note, that you have to move your data to your own data servers/backups within **2 weeks** from your respective measurement time. After the 2-week grace period, the data can be removed from AMIserver without notice due to disk space limitations!

https://ani.aalto.fi 3/4



MEASUREMENT CHECK LIST

Author: Tuomas Tolvanen

Date: 01-10-2020

Else

- EEG settings
- BIOPAC settings
- Eyelink settings
- Stimulus folder
 F Prime setting
- dB levels
- Face camera settings
- fORP settings
- Optical microphone settings

•	Presentation settings Data folder on amiserver	 E-Prime settings 	 PsychoPy settings 	 Psychtoolbox settings
•	Data folder on amiserver	•	•	•

https://ani.aalto.fi 4/4